

Experiments with TOF and steady state (CW) neutrons: **Diffraction Experiments**

Stuart Calder

**Neutron Scattering Division
Oak Ridge National Laboratory**

ORNL is managed by UT-Battelle, LLC for the US Department of Energy


Overview

Experiments with TOF and steady state (CW) neutrons: **Diffraction Experiments**

- General characteristics of Reactor and Spallation sources
 - ORNL has both!
- What this means for diffraction instruments
- Strengths for different experiments
- Consider which source is best suited to your science

Neutron Sources around the world

- Reactor
- Spallation








- Most intense pulsed neutron beam in the world (60Hz, 1.7 MW)




- Highest power research reactor in the world (85 MW)


Neutron sources: Reactor (CW) and Spallation (TOF)



Small $\Delta\lambda$ used, but source on all the time



Each pulse of neutrons contains a broad spectrum of energy (λ)



Pulse of neutrons
~60 times per second


How do you like your neutrons?

CW (HFIR):

Some of the neutrons all of the time?


TOF (SNS):

All of the neutrons some of the time?




What does this mean for
diffraction experiments?

A typical CW diffractometer (HFIR)



$$\lambda = 2d \sin \theta$$


Q range limit is $4\pi/\lambda$



A typical TOF diffractometer (SNS)


De Broglie wavelength: $\lambda = h/mv = ht/mL$

Bragg's law: $\lambda = 2d \sin \theta = \text{constant} * t$



Resolution depends on moderator and t, l, and 2θ

Q range determined by λ_{\min} , λ_{\max} and 2θ



Detector at 2θ measures multiple Q

What are the strengths for diffraction experiments?

A typical CW diffractometer

Strengths

- High flux beam
- Simple beam profile corrections
 - Absorbing, attenuating samples and holders
- Open instruments with variable sample environment
- Tunable resolution and range
- Beam is always on

A typical TOF diffractometer

Strengths

- Widest Q-range
- Resolution high over wide range
- High peak brightness
- Pump probe measurements
- Count rate
 - SNS instruments historically have more detector coverage

Factors to consider for a diffraction experiment

Intensity (CW)

- Higher flux
- Tune intensity with monochromator, collimators

Intensity (TOF)

- Higher peak brightness
- Balance of instrument distance, bandwidth, coverage

Intensity

Q-resolution (TOF)

- Highest resolution
- Complex asymmetric peak shape related to moderator characteristics

Q-resolution

Q-resolution (CW)

- Tunable over Q-range
- Simple, symmetric peak shape function


Q-range (TOF)

- Wide Q coverage
- Q_{\max} can be very high ($>100 \text{ \AA}^{-1}$)
- Good coverage with even limited detectors

Q-range

Q-range (CW)

- Q range limited, but tunable
- Limited Q_{\max} ($\sim 20 \text{ \AA}^{-1}$)
- Need wide/continuous detector coverage



Sample Environment considerations



- An instrument is only useful if the sample is measured under the desired conditions.

<https://neutrons.ornl.gov/sample>

Equipment

- Low Temperature & Magnetic Fields
- High Temperature & Engineering Materials
- High Pressure & Gas Handling
- Special Environments

To only show available items for specific beamline, use this filter.

HB-2A ▾


ID	Sample Space Dia.	Temp Range	Instruments	Associated Resources	Subcategory	Description
ULT-E	40mm	0.03 to 300 K	CG-2, CG-3A, CG-4C, HB-1, HB-1A, HB-2A, HB-2C, HB-3	CRYO-A, CRYO-C, CRYO-O, MAG-B, MAG-E, MAG-G	Insert	Dilution Refrigeration Insert
ULT-G	40mm	0.3 to 300 K	CG-4C, HB-1, HB-1A, HB-2A, HB-2C, HB-3, HB-3A	CRYO-A, CRYO-C, CRYO-N, CRYO-O, MAG-B, MAG-I	Insert	Helium-3 Insert
ULT-H	40mm	0.03 to 300 K	CG-2, CG-4C, HB-1, HB-1A, HB-2A, HB-2C, HB-3, HB-3A	CRYO-A, CRYO-C, CRYO-N, CRYO-O, MAG-B, MAG-E, MAG-G, MAG-I	Insert	Dilution Insert
ULT-J	60mm	0.03 to 300 K	CG-4C, HB-1, HB-1A, HB-2A, HB-2C, HB-3	CRYO-D, CRYO-J, CRYO-K, CRYO-L, CRYO-M	Insert	Dilution Insert
ULT-K	40mm	0.3 to 300 K	CG-2, CG-4C, HB-1, HB-1A, HB-2A	CRYO-A, CRYO-N, CRYO-O, MAG-E, MAG-G, MAG-I	Insert	Helium-3 Insert
ULT-M		0.3 to 300 K	HB-2A, HB-2C, HB-3, HB-3A		Bottom Loading	

Variety of optimized diffraction instruments at SNS/HFIR

HFIR (CW)

- HB-2A
- DEMAND
- IMAGINE
- WAND²
- VERITAS (HB-1A)
- HIDRA
- GP-SANS


- HB-1
- HB-3
- CTAX



SNS (TOF)



- POWGEN
- NOMAD
- TOPAZ
- CORELLI
- MANDI
- SNAP
- VULCAN
- EQ-SANS

- HYSPEC
- ARCS
- SEQUOIA
- CNCS



Is diffraction best at CW or TOF instruments?

- It depends on the question to be answered in your science!



Let's consider some common questions diffraction can answer and match to strengths of sources

- Measure nuclear structure
- Measure magnetic structure
- Measure disorder in structure
- ... Lots more cases for you to consider!

Measure a crystalline structure

Need


- To collect lots of reflections

Instrument should have

- High Q-resolution (**TOF/CW**)
- Large Q-range (**TOF**)
- High Q (**TOF**)

• **TOF is typically best, especially for powders.**

• **Single crystal: TOF for largest unit cells, but CW can work well.**





Measure the crystal structure

Instruments at HFIR/SNS

- Powder samples
 - **POWGEN (TOF)**: High resolution and Q coverage
 - **NOMAD (TOF)**: High flux and Q coverage
 - **HB-2A (CW)**: Smaller unit cell, complex SE, absorbing samples.

- Single crystals
 - **DEMAND (CW)**: Smaller unit cell inorganic materials
 - **TOPAZ (TOF)**: High resolution and coverage for inorganic/organic and larger structures
 - **IMAGINE (CW)**: Quasi-Laue for macromolecules
 - **MANDI (TOF)**: Protein crystallography



DEMAND (10^4 \AA^3)
TOPAZ (10^5 \AA^3)
IMAGINE (10^6 \AA^3)
MaNDi (10^7 \AA^3)

Measure the Magnetic structure

Need

- To measure where magnetic scattering is strongest.
- Good signal to noise.
- Variety of sample conditions.

Instrument should have

- Low Q (**CW/TOF**)
- High intensity (**CW**)
- Low background (**CW/TOF**)
- Low T and magnets (**CW**)

- **CW is typically best, but TOF has ever increasing options.**

Short-range order

Specialized SE

Polarization

Absorbing, H

Specific Q region

Small sample

TOF/CW
Consider low Q or mPDF options

CW/TOF

CW

- CW easier.
- Consider TAS.
- TOF feasible





- TOF/CW
- TOF: If need both high/low CW:
- One region

TOF/CW

Measure the magnetic structure

Instruments at HFIR/SNS

- Powder samples
 - **HB-2A (CW)**: Access low Q. Low Temperature and magnets. Polarization.
 - **POWGEN (TOF)**: High resolution and Q coverage.
- Single crystal
 - **DEMAND (CW)**: Low Q coverage and variety of sample environments. Polarization.
 - **TOPAZ (TOF)**: Wide coverage in Q. New 5 K option.
 - **CORELLI (TOF)**: Diffuse scattering. Variety of sample environments.
 - **GP-SANS (CW)**: Very low Q for large spin textures (e.g. Skyrmions).
- Both Powder and single crystal
 - **WAND² (CW)**: High flux → Long range and Diffuse signals
 - **HB-1A (CW)**: Excellent signal-to-noise for weak signal



Disordered materials (PDF)


Need

- To access high Q and broad scattering

Instrument should have

- Wide Q-range (TOF)
- High Q_{max} (TOF)
- High intensity (TOF/CW)




• TOF is typically best, especially for powders.
• But CW can work well.




Disordered material (PDF)

Instruments at HFIR/SNS


- Powder
 - **NOMAD (TOF)**: Dedicated total scattering beamline.
 - **POWGEN (TOF)**: Longer counting, but better resolution if needed.
 - **HB2A and WAND² (CW)**: Potential for PDF and mPDF
- Single crystal
 - **CORELLI (TOF)**: Dedicated diffuse scattering beamline
 - **TOPAZ (TOF)**: Large reciprocal space coverage
 - **WAND² (CW)**: High flux and variety of sample environments



Many more diffraction experiments



Engineering diffraction
Measure strain/stress




**VULCAN
(TOF)**

- In situ loading studies of crystalline/amorphous materials at high temperatures: phase transformation, fatigue damage, creep behaviors, and other deformation mechanisms in nanostructured materials, piezoelectric and shape-memory alloys.

**HIDRA
(CW)**

- Optimized for strain measurement and determination of residual stress in engineering materials.
- Spatial resolution at a fraction of a millimeter is possible

Many more diffraction experiments



SNAP
(TOF)

- Dedicated instrument for high pressure >100GPa.
- Powders and single crystals.
- Wide Q-coverage
- PDF options

High pressure
diffraction

WAND²,
DEMAND,
HB-2A,
TAS (CW)

- Variety of options for pressure measurements.



Review articles for the Diffraction Suite

REVIEW OF SCIENTIFIC INSTRUMENTS 89, 092802 (2018)

A suite-level review of the neutron single-crystal diffraction instruments at Oak Ridge National Laboratory

L. Coates,^{1,a)} H. B. Cao,¹ B. C. Chakoumakos,¹ M. D. Frontzek,¹ C. Hoffmann,¹ A. Y. Kovalevsky,¹ Y. Liu,¹ F. Meilleur,^{1,2} A. M. dos Santos,¹ D. A. A. Myles,¹ X. P. Wang,¹ and F. Ye¹

¹Neutron Scattering Division, Oak Ridge National Laboratory, 1 Bethel Valley Road, Oak Ridge, Tennessee 37831, USA

²Department of Molecular and Structural Biochemistry, North Carolina State University, Raleigh, North Carolina 27695, USA

(Received 26 March 2018; accepted 9 July 2018; published online 24 September 2018)

The nascent suite of single-crystal neutron diffractometers at the Oak Ridge National Laboratory has no equal at any other neutron scattering facility worldwide and offers the potential to re-assert single-crystal diffraction using neutrons as a significant tool to study nuclear and magnetic structures of small unit cell crystals, nuclear structures of macromolecules, and diffuse scattering. Signature applications and features of single-crystal neutron diffraction are high resolution nuclear structure analysis, magnetic structure and spin density determinations, contrast variation (particularly D₂O/H₂O) for nuclear structural studies, lack of radiation damage when using crystals of biological molecules such as proteins, and the fidelity to measure nuclear and magnetic diffuse scattering with elastic discrimination. © 2018 Author(s). All article content, except where otherwise noted, is licensed under a Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>). <https://doi.org/10.1063/1.5030896>

REVIEW OF SCIENTIFIC INSTRUMENTS 89, 092701 (2018)

A suite-level review of the neutron powder diffraction instruments at Oak Ridge National Laboratory

S. Calder,^{1,a)} K. An,¹ R. Boehler,^{1,2} C. R. Dela Cruz,¹ M. D. Frontzek,¹ M. Guthrie,^{3,4} B. Haberl,¹ A. Huq,¹ S. A. J. Kimber,¹ J. Liu,¹ J. J. Molaison,¹ J. Neufeind,¹ K. Page,¹ A. M. dos Santos,¹ K. M. Taddei,¹ C. Tulk,¹ and M. G. Tucker¹

¹Neutron Scattering Division, Oak Ridge National Laboratory, 1 Bethel Valley Rd., Oak Ridge, Tennessee 37831, USA

²Geophysical Laboratory, Carnegie Institution of Washington, Washington, District of Columbia 20015, USA

³European Spallation Source, Lund 221 00, Sweden

⁴University of Edinburgh, Edinburgh EH8 9YL, United Kingdom

(Received 6 April 2018; accepted 19 July 2018; published online 28 September 2018)

The suite of neutron powder diffractometers at Oak Ridge National Laboratory (ORNL) utilizes the distinct characteristics of the Spallation Neutron Source and High Flux Isotope Reactor to enable the measurements of powder samples over an unparalleled regime at a single laboratory. Full refinements over large Q ranges, total scattering methods, fast measurements under changing conditions, and a wide array of sample environments are available. This article provides a brief overview of each powder instrument at ORNL and details the complementarity across the suite. Future directions for the powder suite, including upgrades and new instruments, are also discussed. © 2018 Author(s). All article content, except where otherwise noted, is licensed under a Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>). <https://doi.org/10.1063/1.5033906>

Conclusions

- ORNL is unique in having world class CW and TOF sources.
- HFIR and SNS diffraction instruments have different strengths.
- Choose the best instrument for your experiment based on your science.

Lots of good options for diffraction at HFIR/SNS!





Stuart Calder
POWDER

ASK AN INSTRUMENT SCIENTIST!!!


<https://neutrons.ornl.gov/suites/diffraction>



Arianna Minelli
Instrument Scientist
CORELLI



Andrey
Kovalevsky
IMAGINE, MANDI



Cheng Li
NOMAD,
POWGEN



Clarina dela Cruz
POWDER



Keith Taddei
POWDER,
WAND²



Joerg C.
Neufeind
NOMAD



Dean Myles
IMAGINE, MANDI



Yan Chen
VULCAN



Thomas Proffen
POWGEN



Jue Liu
NOMAD



Qiang Zhang
POWGEN



Alicia Manjón
Sanz
POWGEN




Ke An
VULCAN



Xiaoping Wang
TOPAZ



Matthias Frontzek
WAND²




Huibo Cao
DEMAND



Yan Wu
DEMAND, WAND²



Bryan
Chakoumakos
DEMAND



Feng Ye
CORELLI



Flora Meilleur
IMAGINE, MANDI




Chris Tulk
SNAP



Jeffrey Bunn
HIDRA



Andrew Payzant
HIDRA



Chris Fancher
HIDRA



António M. dos
Santos
SNAP



Christina
Hoffmann
CORELLI, TOPAZ

Questions?

Lots of good options for diffraction at HFIR/SNS!



NXS Lecture - Diffraction & Spectroscopy at TOF vs Continuous Sources

