


DEMAND

Dimensional Extreme Magnetic Neutron Diffractometer

The Dimensional Extreme Magnetic Neutron Diffractometer (DEMAND) has two modes of operation. Four-circle mode with the circle goniometer can operate with a closed-cycle-refrigerator (4–800 K) that allows a pressure cell (0–10 GPa), or a permanent magnet (0–0.9 T), and electric field. Two-axis mode allows extreme sample equipment cryomagnet (0–6 T), ^3He and dilution refrigerators (0.05–300 K). Both modes can have an unpolarized and polarized neutron beam that can be switched by automated elevator device. The current detector has three Anger camera modules stacking vertically to cover 16° (hori) \times 48° (vert). The detector sits on the 2Θ rotation arm to cover the scattering angle range of $3^\circ < 2\Theta < 155^\circ$. A multilayer-[110]-wafer silicon monochromator with the reflection from HHL planes ensures sharp diffraction peaks in specified ranges of detector angles by control of the horizontal radius of curvature. Any HHL planes can be set in Bragg position, but only the (331), (220) with (440), and (111) with (333) reflections are of practical interest. For the fixed monochromator angle of 47.5° , these reflections provide principal incident wavelengths of 1.005 Å, 1.546 Å, and 2.541 Å, respectively. A PC-based LabView system provides user-friendly diffractometer control and data acquisition. The beam size is 6.3 mm in diameter, and the minimum measured crystal size is 0.02 mm^3 . The maximum crystal dimension is



usually limited to 5 mm. The flux on the sample can be up to $2.2 \times 10^7 \text{ n/cm}^2/\text{s}$. The horizontal bending of the monochromator can be changed to optimize the Q-resolution or flux depending upon the needs of the measurement.

APPLICATIONS

The HB-3A DEMAND has the mission to explore nuclear and magnetic structures as a function of temperature, pressure, magnetic field, and electric field. The instrument is particularly suitable for studying phase transitions and accompanying structure changes, as well as measuring order parameters and exploring the phase diagram. It also is suitable for a wide range of small-unit-cell crystallography studies, including superlattice structures and atomic anharmonicity. Users have researched problems in physics, materials science, chemistry, and mineralogy. Recent topics can be found in the publication and science highlight pages.

SPECIFICATIONS

Beam spectrum	Thermal
Monochromator	Double focusing silicon
Monochromator angle	47.5°
Incident wavelengths	1.005 Å (331), 1.546 Å (220), 2.541 Å (111)
Four-circle mode	Huber, full circle, with 4 – 800 K CCR $2\Theta < 95^\circ$
Two-axis mode	$-181^\circ < \omega < 181^\circ$ Out-of-scattering plane coverage $-8^\circ < \psi < 40^\circ$ $2\Theta < 95^\circ$
Detector	2D Anger camera
Crystal size requirement	$>0.1 \text{ mm}^3$, maximum crystal dimension 5 mm
Flux at sample	$2.2 \times 10^7 \text{ n/cm}^2/\text{s}$
Polarized neutron diffraction	Polarizer: S-bender super mirror ^3He Polarizer

26-G023702/cpp April 2026

For more information, contact

Huibo Cao, caoh@ornl.gov, 865.686.2608

Bryan Chakoumakos, chakoumakobc@ornl.gov, 865.621.0227

Yan Wu, wuy1@ornl.gov, 865.341.0563

neutrons.ornl.gov/demand

Managed by UT-Battelle LLC
for the US Department of Energy